



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/828,359	04/20/2004	Tom Westberg	F-5495 DIV (0360-0172.01)	4891
44926 7590 06/14/2007 BAXTER HEALTHCARE CORPORATION ONE BAXTER PARKWAY DF2-2E DEERFIELD, IL 60015			EXAMINER WIEST, PHILIP R	
			ART UNIT 3761	PAPER NUMBER
			MAIL DATE 06/14/2007	DELIVERY MODE PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary	Application No. 10/828,359	Applicant(s) WESTBERG ET AL.	
	Examiner Phil Wiest	Art Unit 3761	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☐ Responsive to communication(s) filed on 17 April 2004.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-6 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1-6 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☒ The drawing(s) filed on 17 April 2004 is/are: a) ☒ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|---|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| 3) <input checked="" type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08)
Paper No(s)/Mail Date <u>4/17/04, 6/10/04</u> . | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION***Double Patenting***

1. The nonstatutory double patenting rejection is based on a judicially created doctrine grounded in public policy (a policy reflected in the statute) so as to prevent the unjustified or improper timewise extension of the "right to exclude" granted by a patent and to prevent possible harassment by multiple assignees. A nonstatutory obviousness-type double patenting rejection is appropriate where the conflicting claims are not identical, but at least one examined application claim is not patentably distinct from the reference claim(s) because the examined application claim is either anticipated by, or would have been obvious over, the reference claim(s). See, e.g., *In re Berg*, 140 F.3d 1428, 46 USPQ2d 1226 (Fed. Cir. 1998); *In re Goodman*, 11 F.3d 1046, 29 USPQ2d 2010 (Fed. Cir. 1993); *In re Longi*, 759 F.2d 887, 225 USPQ 645 (Fed. Cir. 1985); *In re Van Ornum*, 686 F.2d 937, 214 USPQ 761 (CCPA 1982); *In re Vogel*, 422 F.2d 438, 164 USPQ 619 (CCPA 1970); and *In re Thorington*, 418 F.2d 528, 163 USPQ 644 (CCPA 1969).

A timely filed terminal disclaimer in compliance with 37 CFR 1.321(c) or 1.321(d) may be used to overcome an actual or provisional rejection based on a nonstatutory double patenting ground provided the conflicting application or patent either is shown to be commonly owned with this application, or claims an invention made as a result of activities undertaken within the scope of a joint research agreement.

Effective January 1, 1994, a registered attorney or agent of record may sign a terminal disclaimer. A terminal disclaimer signed by the assignee must fully comply with 37 CFR 3.73(b).

2. Claims 1-3 and 6 are rejected on the ground of nonstatutory obviousness-type double patenting as being unpatentable over claims 1, 3, 4, 7, and 8 of U.S. Patent No. 7,041,076. Although the conflicting claims are not identical, they are not patentably distinct from each other because both applications disclose a blood processing system comprising a leukocyte filter, a blood source, a blood collection container that receives blood from a blood separation device, a plurality of pump stations, and a fluid pressure-activator that operates the pump stations in multiple modes.

Claim Rejections - 35 USC § 103

3. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

4. The factual inquiries set forth in *Graham v. John Deere Co.*, 383 U.S. 1, 148 USPQ 459 (1966), that are applied for establishing a background for determining obviousness under 35 U.S.C. 103(a) are summarized as follows:

1. Determining the scope and contents of the prior art.
2. Ascertaining the differences between the prior art and the claims at issue.
3. Resolving the level of ordinary skill in the pertinent art.
4. Considering objective evidence present in the application indicating obviousness or nonobviousness.

5. This application currently names joint inventors. In considering patentability of the claims under 35 U.S.C. 103(a), the examiner presumes that the subject matter of the various claims was commonly owned at the time any inventions covered therein were made absent any evidence to the contrary. Applicant is advised of the obligation under 37 CFR 1.56 to point out the inventor and invention dates of each claim that was not commonly owned at the time a later invention was made in order for the examiner to consider the applicability of 35 U.S.C. 103(c) and potential 35 U.S.C. 102(e), (f) or (g) prior art under 35 U.S.C. 103(a).

6. Claims 1-6 are rejected under 35 U.S.C. 103(a) as being unpatentable over Dennehey et al. (US 5,482,440) in view of Lundback (US 4,750,868).

7. Dennehey et al. (hereafter Dennehey) discloses a blood processing system comprising a leukocyte filter 82, a plurality of fluid pressure actuators, a plurality of blood collection containers 20 that receive blood from a blood separation device, as per Claims 2 and 3. The pressure actuators operate to apply fluid pressure strokes in tandem to a first and second pump station to convey fluid therethrough (Column 23, Line 61 through Column 26, Line 67). The pressure actuation system comprises a control function that has two alternating flow modes (draw cycle and return cycle) that ensure a continuous flow of fluid through the cassette (Column 18, Lines 4-8). Regarding Claim 4, Dennehey discloses that the plurality of pump stations apply a positive and negative pneumatic pressure (Column 25, Lines 45-53). Regarding Claim 4, Pneumatic pumps utilize air, which is a fluid. Dennehey also discloses a blood collection container 16 located downstream of the filter 82 that receives blood after it passes through the filter, as per Claim 6 (see Figure 2). Dennehey, however, does not specifically disclose that the fluid pressure actuator synchronizes the pump strokes such that fluid flow to the filter is pulsatile.

8. Lundback discloses a pumping system and method that uses a pneumatic driving means to provide a pulsatile outflow and a continuous inflow. The system of Lundback comprises two pump stations, A and V, which are in communication between the fluid inlet and outlet (see FIG 3A). The method disclosed by Lundback includes a pumping stroke and a return stroke that operate in succession to one another in order to create a continuous inflow and a pulsatile outflow (see columns 6-7). Since the Dennehey device meets the needs of various pumping functions, it would have been obvious to one of

Art Unit: 3761

ordinary skill in the art at the time of invention to combine the blood processing system of Dennehey with the pumping method of Lundback in order to provide a continuous inflow and pulsatile outflow in order to provide a pressure-sensitive pumping system, as taught by Lundback.

Conclusion

9. Any inquiry concerning this communication or earlier communications from the examiner should be directed to Phil Wiest whose telephone number is (571) 272-3235. The examiner can normally be reached on 8:30am-5pm EST.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Tatyana Zalukaeva can be reached on (571) 272-1115. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

Application/Control Number: 10/828,359

Page 6

Art Unit: 3761

PRW

6/7/07

TATYANA ZALUKAEVA
SUPERVISORY PRIMARY EXAMINER

A handwritten signature in black ink, appearing to read 'Tatyana', written in a cursive style.